

Certificate of Compliance

Product : AC Adapter
Applicant : Pihong Technology Co., Ltd.
Rating(s) : Input: AC100-240V, 50-60Hz, 0.6-0.25A
Output: (1) 12Vdc, 1.5A; (2) 24Vdc, 0.75A
Model/Type : (1) PMA18A-120, PMA18E-120,
PMA18K-120, PMA18L-120,
PMA18S-120
(2) PMA18A-240, PMA18E-240,
PMA18K-240
Test Report No. : H180822-05-A0
Standard(s) Applied : IEC 60601-1-11: 2015 (Second Edition);
EN 60601-1-11: 2015

We hereby declare that the above equipment has been tested in our facility with the listed standard and found in compliance with the harmonized Standards of Medical Device Directive 93/42/EEC




Tony Hsu / Manager

2018-12-14



Prodigy Technology Consultant Co.Ltd.

No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan CHINESE TAIPEI
TEL: 886-2-2603-7288/FAX: 886-2-2602-0908