

Technical Compliance Statement

EMC Test Report

For the following information

Ref. File No.: C1M1809052

Product : AC Adapter
Model Number : (1)PMA18E-120 (2)PMA18K-120 (3)PMA18L-120
(4)PMA18A-120 (5)PMA18A-240 (6)PMA18E-240
(7)PMA18K-240 (8)PMA18S-120



Brand Name :
Applicant : Phihong Technology Co., Ltd.
Manufacturer : Phihong Technology Co., Ltd.
Standards :

EN 60601-1-2:2015 (IEC 60601-1-2: 2014)

Emission: CISPR 11:2009 +A1:2010, Class B
IEC 61000-3-2:2014 and IEC 61000-3-3:2013

Immunity: IEC 61000-4-2:2008, IEC 61000-4-3:2010, IEC 61000-4-4:2012,
IEC 61000-4-5:2014 +A1:2017, IEC 61000-4-6:2013,
IEC 61000-4-8:2009, IEC 61000-4-11:2004 +A1:2017

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council EMC directive 93/42/EEC. The test data and results are issued on the EMC test report no. **EM-E180554**.

Signature

Alex Deng/Deputy Manager

Date: 2018. 10. 25

Test Laboratory:
Audix Technology Corporation, EMC Department
TAF Accreditation No.: 1724
Web Site: www.audixtech.com

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.