

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Open Frame Medical Power Supply
Name and address of the applicant	Phihong Technology Co., Ltd. No. 568 Fuxing 3rd Rd. Guishan District, Taoyuan City 333 Taiwan
Name and address of the manufacturer	Phihong Technology Co., Ltd. No. 568 Fuxing 3rd Rd. Guishan District, Taoyuan City 333 Taiwan
Name and address of the factory	Boayang Electronics Co., Ltd. Di Feng Gong Ye Qu 2 Hao Xiasha Liuwu Village Shipai Town, Dong Guan City, 523350 Guang Dong Province, China
Ratings and principal characteristics	AC 100-240V; 50-60Hz; 1.6-0.7A; Class II; for output ratings refer to test report page 6
Trademark (if any)	PHIHONG
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	PML065-xxxxxy (xxx = 120, 150, 180, 190, 240 or 480 to denote different output rating; yyy = any alphanumeric character or blank for marketing purpose only)
Additional information (if necessary may also be reported on page 2)	This power supply has been evaluated as a Class II and continuously operating equipment. An additional evaluation shall be made if the power supply is intended for use in Class I; for details see test report
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1 See Test Report for National Differences
As shown in the Test Report Ref. No. which forms part of this Certificate	50307111 001

This CB Test Certificate is issued by the National Certification Body



**TÜVRheinland**®

TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web: www.tuv.com



Date: 30.10.2019

Signature:

Dipl.-Ing. (FH) A. Klinker