

Phihong Technology Co., Ltd.
Ms. Cathy Weng Safety Engineer
Safety Compliance Section
No. 568, Fuxing 3rd Rd.,
Guishan District, Taoyuan City 333
Taiwan

Date : 22.10.2019
Our ref. : ZTW1
Your ref.: TH/SPC- 1908006/SPC

Ref : TA License

Type of Equipment : Open Frame Medical Power Supply
Model Designation : See Certificate
Certificate No. : TA 50448706 0001
Report No. : 50059740 002

Dear Ms. Cathy Weng,

The above specified technical equipment has been tested and found to be in accordance with the relevant requirements.

Enclosed you will find the certificate of approval.

If cancellation of the certificate is submitted by November 15 in a given year, no fee will be charged for the following year.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Please contact our office for approval of your new products.

With kind regards,

Certification Body



Dipl.-Ing. (FH) A. Klinker

cc: Superior Product Consulting, Inc.

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
<http://www.tuv.com>

TAICHUNG BRANCH:
No. 9, Lane 36, Sec. 3, Minsheng Rd.,
Daya Dist., Taichung City 428,
Taiwan, R. O. C.
Tel. (04) 2521-8888
Fax (04) 2521-8899

KAOSHIUNG BRANCH:
27F-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

Certificate



Certificate no.

TA 50448706 01

License Holder:

Phihong Technology Co., Ltd.
No. 568, Fuxing 3rd Rd.,
Guishan District, Taoyuan City 333
Taiwan

Manufacturing Plant:

030-0000427402

Test report no.:

ZTW1- 50059740 002

Client Reference: TH/SPC- 1908006/SPC

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product:

(Open Frame Medical Power Supply)

License Fee - Units

Type Designation	: PML065-xxxxyy (PHIHONG)	5
xxx stands for	: 120, 150, 180, 190, 240, 480	1
yyy stands for	: any alphanumeric character or blank	1
Rated Voltage	: AC 100-240V, 50-60Hz	
Rated Current	: 1.6-0.7A	
Outputs	: see Appendix	
max. Ambient Temperature:	50°C	
Protection Class	: II	

Remark: 1. This certificate is valid for maximum 5 years.
2. This power supply has been evaluated as a Class II and continuously operating equipment. An additional evaluation shall be made if the power supply is intended for use in Class I equipment.

7

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419073981

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

22/10/2019


TÜV Rheinland Group



Appendix to TÜV approved Certificate No.: TA 50448706

Certified Product : Open Frame Medical Power Supply

Report Number : 50059740 002

Type Designation: PML065-xxxxyy (where xxx = 120,150,180,190,240 or 480, yyy = any alphanumeric character or blank) ()

Output rating table:

Pihong	Output Voltage(Vdc)	Output Current(A)
PML065-120yyy	12V	5.0A
PML065-150yyy	15V	4.34A
PML065-180yyy	18V	3.62A
PML065-190yyy	19V	3.43A
PML065-240yyy	24V	2.71A
PML065-480yyy	48V	1.36A

Date: Oct. 22, 2019



Certification Body

Dipl.-Ing. (FH) A. Klinker